



DryView Chroma Imaging System DICOM Conformance Statement

Feb 08, 2011

Document # 8J9581

Revision 1.0

Carestream Health reserves the right to change any part of this document without prior notice. Federal Copyright law, with all rights reserved, protects this publication.

Note: Publication of a DICOM Conformance Statement does not indicate product availability in all countries. Please contact your Carestream Health representative, or Carestream Health in your country, for information on availability of a specific product.

Contents

1	Introduction	1
1.1	Executive Overview	1
1.2	Scope and Field of Application	1
1.3	Important Considerations for the Reader	1
1.4	Accessing this Conformance Statement on the World Wide Web	2
1.5	Definitions, Acronyms, Abbreviations	2
2	Implementation Model	4
2.1	Functional Definitions	4
2.2	Sequencing of Real-World Activities	4
3	Application Entity Specifications.....	5
3.1	Association Establishment Policies.....	5
3.1.1	General.....	5
3.1.2	Number of Associations.....	5
3.1.3	Asynchronous Nature.....	5
3.1.4	Implementation Identifying Information.....	5
3.2	Association Acceptance Policy	6
3.2.1	Associated Real-World Activity	6
3.2.2	Presentation Context Table	8
3.2.3	SOP Specific Conformance.....	9
3.3	Basic Print Management Meta SOP Class	9
3.3.1	Basic Film Session SOP Class	10
3.3.2	Basic Film Box SOP Class.....	12
3.3.3	Basic Grayscale Image Box SOP Class	18
3.3.4	Basic Color Image Box SOP Class	20
3.3.5	Printer SOP Class.....	23
3.4	Basic Annotation Box SOP Class.....	25
3.4.1	DIMSE Service N-SET	25
3.4.2	Annotation N-Set Status Code	26
3.5	Presentation LUT SOP Class.....	26
3.5.1	DIMSE Service N-CREATE.....	26
3.5.2	Presentation LUT N-Create Status Code.	27
3.5.3	DIMSE Service N-DELETE.....	27
3.5.4	Presentation LUT N-Delete Status Code	27
4	Communication Profiles	28
4.1	Supported Communications Stacks	28
4.2	Physical Media.....	28
5	Extensions/Specializations/Privatizations	28
6	Configuration.....	28
7	Support of Extended Character Sets.....	29
8	Error Handling	29
	Annex A - Configuration Information.....	30
	Annex B - Custom Formats	33

Annex C - Unsupported Film Types	34
Annex D – Printer Specifications	36
D.1 Medium Size	36
D.2 Printer Capabilities	37

1 Introduction

1.1 Executive Overview

This document covers the DryView Chroma Imaging System (referred to as “Imager” henceforth).

The following DICOM SOP Classes are supported:

SOP Class Name	SOP Class UID	Service Class Role
Verification SOP Class	1.2.840.10008.1.1	SCP
Basic Grayscale Print Management Meta SOP Class	1.2.840.10008.5.1.1.9	SCP
Basic Color Print Management Meta SOP Class	1.2.840.10008.5.1.1.18	SCP
Basic Annotation Box SOP Class	1.2.840.10008.5.1.1.15	SCP
Presentation LUT SOP Class	1.2.840.10008.5.1.1.23	SCP

1.2 Scope and Field of Application

This document describes the DICOM functionality of the Imager. The Imager captures exam images from DICOM sources and prints the images. The Imager acts as a DICOM Service Class Provider (SCP). The Imager performs transactions over a TCP/IP network via the DICOM messages exchange protocol.

1.3 Important Considerations for the Reader

This DICOM Conformance Statement is written in such a way that assumes the reader has a basic understanding of DICOM. This DICOM Conformance Statement by itself is not sufficient to guarantee successful connectivity between the Imager and equipment from other vendors. The following considerations should be made:

- The integration of equipment from different vendors (including Carestream Health) goes beyond the scope of the DICOM 3.0 standard and the DICOM Conformance Statements from Carestream Health and other vendors. It is the responsibility of the user (or user’s agent) to assess the application requirements and to design a solution that integrates Carestream Health equipment with equipment from other vendors.
- When the comparison of this DICOM Conformance Statement with a DICOM Conformance Statement from another vendor indicates that connectivity should be possible, it is the responsibility of the user (or user’s agent) to verify this by carrying out validation tests and to check whether all required functionality is met.
- Carestream Health reserves the right to make changes to the Imager architecture described in this document. The user (or user’s agent) should ensure that any equipment connected via DICOM to Carestream Health equipment also follows the future evolution of the DICOM 3.0 standard. Failure to do so may result in (partial) loss of connectivity.

- For all DICOM attributes of type M (shown in the column of SCP Usage), the SCU must send a valid value within the published range. A missing attribute would result in returning an error of Missing Attribute, an invalid value would result in returning an error of Invalid Value and an Out of Range value would result in returning an error of Out of Range. It is up to the SCU to retry with a new value or to abort the association.
- For all DICOM attributes of type U (shown in the column of SCP Usage), the Default Value or the Configured Value (if it exists) is used for all cases of Missing Attribute, Invalid Value or Out of Range. No error would be generated from this type. The imager will send back the corrected value (i.e. the value being used) to the SCU. It is up to the SCU to accept, to decline (abort) or to resend a new value at this point.
- Attributes received at the Image Box level will override the same attribute received at the Film Box level for a particular image with the exception of the Tonescaling method specified in configuration information. Only one Tonescaling method is supported in one film box (Curve Shape or Perception LUT), the image box Perception LUT cannot override the film Box Curve Shape.

1.4 Accessing this Conformance Statement on the World Wide Web

As the Imager product changes, changes to this DICOM Conformance Statement are inevitable. To obtain the most recent revision of this DICOM Conformance Statement, access the following URL:

<http://www.carestream.com/dicom.html>

1.5 Definitions, Acronyms, Abbreviations


The following symbols and abbreviations are used in this document.

Acronym/Abbreviation	Definition
AE	Application Entity
ASCII	American Standard Code for Information Interchange
DICOM	Digital Imaging and Communications in Medicine
DIMSE	DICOM Message Service Element
IOD	Information Object Definition.
ISO	International Standards Organization
LUT	Look-up Table
M	Indicate Mandatory when used in the SCP Usage columns.
MC	Indicate Mandatory with Conditions when used in the SCP Usage columns.
PDU	Protocol Data Unit
PLUT	Presentation Look-Up Table
SCP	Service Class Provider
SCU	Service Class User
SOP	Service-Object Pair

Acronym/Abbreviation	Definition
TCP/IP	Transmission Control Protocol/Internet Protocol
TFT/ULUT	Transfer Function Table
U	Indicate User choice (Not Mandatory) when used in the SCP Usage columns.
UID	Unique Identifier

2 Implementation Model

This implementation model uses the DICOM Basic Print Management Meta SOP Class to receive studies for the Imager. Associations from multiple SCUs are supported.



2.1 Functional Definitions

The Imager acquires images from the connected device(s), temporarily stores them on disk, and then formats and prints them to film.

2.2 Sequencing of Real-World Activities

The Imager prints images to film upon request from the SCU (i.e. N-ACTION-RQ). It operates as required to meet the definition of the Print Management Service Class.

3 Application Entity Specifications

The Imager provides Standard Conformance to the following SOP Classes as an SCP.

SOP Class Name	SOP Class UID
Verification SOP Class	1.2.840.10008.1.1
Basic Grayscale Print Management Meta SOP Class	1.2.840.10008.5.1.1.9
Basic Color Print Management Meta SOP Class	1.2.840.10008.5.1.1.18
Basic Annotation Box SOP Class	1.2.840.10008.5.1.1.15
Presentation LUT SOP Class	1.2.840.10008.5.1.1.23

3.1 Association Establishment Policies

3.1.1 General

3.1.1.1 Delivery - Basic Grayscale/Color Print Management SCP

Chroma's maximum PDU size is 128 Kbytes.

Chroma is a Grayscale and Color imager. The Basic Color Print jobs will be printed on paper and Basic Grayscale Print jobs will be printed on film.

3.1.2 Number of Associations

Imager will accept by default maximum of 2 simultaneous associations, but this is configurable and can be modified. If an attempt is made to open more than the number of configured simultaneous associations, the Imager will reject the additional associations (A-ASSOCIATE-RJ). Note: The A-ASSOCIATE-RJ will be transient and not permanent...See 3.2.1.4, Association Rejection.

3.1.3 Asynchronous Nature

The Imager allows up to 1 invoked and 1 performed operation on an Association (it is synchronous, e.g. the SCU/SCP can send only 1 Request and must wait for the corresponding Response before sending the next Request).

3.1.4 Implementation Identifying Information

The Imager provides the Implementation Class UID of "1.2.840.113564.3.3.3"

The implementation version name attribute is of the form of "DVChroma_{yy}v_{xxx}" where yy is the Released Year and xxx is the Version Number. For example: DVChroma11v1.0 stands for DryView Chroma Imager software, released in 2011, with 1.0 as the version.

The Called AE Title used by the Imager is the hostname.

3.2 Association Acceptance Policy

3.2.1 Associated Real-World Activity

3.2.1.1 Delivery - Basic Grayscale/Color Print Management SCP

The Imager accepts Associations for the purpose of acquiring images and printing them on a local printer.

When an association has been established:

- The SCU can request the Imager to create or set Film Sessions, Film Boxes, set Image boxes, Annotation boxes, Presentation LUT and request Printer Status
- The SCU can request to change the attributes that are allowed for these boxes.

The port number is configurable and the default setting is 5040.

3.2.1.2 Application Entity Titles (AE_TITLE)

The Called AE Title for the Imager is “DVChroma”.

The Called AE Title may be used to select behavior, which is unique to *Carestream Health* legacy products for backwards compatibility:

- Suffix “/C” option of the AE Title
Curve shape tone scaling values are usually interpreted in standard 0-999 range. However, to support Carestream Health legacy devices, the Imager’s Print Server can be configured to accept the curve shape in the 0-690 range. If the Called AE Title contains a suffix “/C”, the 0-690 range shall be used.
- Suffix “/1..5” option of the AE Title is ignored.
The Chroma Imager has only one single output tray. The Suffix “/1..5” is for supporting legacy products only. Therefore, this suffix has no effect.
- ◆ NER_ option of the AE Title is ignored and will not provide status changes in the Imager as they occur. It is necessary to turn on the N-Event-Report from the Imager’s User Interface.

3.2.1.3 Association Negotiation

If the association is accepted, the list of requested Presentation Context items is returned with each item marked as accepted or rejected with the Result/Reason field containing the values specified in PS 3.8 Table 9-18 of the DICOM Standard, Release PS 3.1-2004.

3.2.1.4 Association Rejection

If the association is rejected, the Result, Source, and Reason/Diagnostic fields in the response message contain the values show below:

Condition	Result	Source	Reason/Diagnostic
Limit on simultaneous associations exceeded	2 – Rejected Transient	3 - DICOM UL Service Provider (Presentation)	3 – local limit exceeded
The IP Connection could not established.	2 – Rejected Transient	3 – DICOM UL Service Provider (Presentation)	1 – temporary congestion
The destination printer is recognized, but not installed.	1 - Rejected Permanent	1 - DICOM UL User	1 – no reason given
No Implementation UID	1 - Rejected Permanent	2 – DICOM UL Service Provider (ACSE)	1 – no reason given
No Application Context Name	1 – Rejected Permanent	2 – DICOM UL Service Provider (ACSE)	1 – no reason given
DICOM protocol version is not supported	1 – Rejected Permanent	2 – DICOM UL Service Provider (ACSE)	2 – protocol version not supported
No Presentation Context items given.	1 – Rejected Permanent	2 – DICOM UL Service provider (ACSE)	1 – no reason given
No presentation context items accepted	1 – Rejected Permanent	1 - DICOM UL User	1 – no reason given

3.2.2 Presentation Context Table

The Imager accepts the Presentation Contexts shown below.

Presentation Context Table

Abstract Syntax		Transfer Syntax		Role	Ext.
Name	UID	Name List	UID		Negot
Verification	1.2.840.10008.1.1	DICOM Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		DICOM Explicit VR Little Endian	1.2.840.10008.1.2.1		
Basic Grayscale Print Management	1.2.840.10008.5.1.1.9	DICOM Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		DICOM Explicit VR Little Endian	1.2.840.10008.1.2.1		
Basic Color Print Management	1.2.840.10008.5.1.1.18	DICOM Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		DICOM Explicit VR Little Endian	1.2.840.10008.1.2.1		
Basic Annotation Box	1.2.840.10008.5.1.1.15	DICOM Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		DICOM Explicit VR Little Endian	1.2.840.10008.1.2.1		
Presentation LUT	1.2.840.10008.5.1.1.23	DICOM Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		DICOM Explicit VR Little Endian	1.2.840.10008.1.2.1		

3.2.3 SOP Specific Conformance

3.2.3.1 Verification

The Imager provides standard conformance to the DICOM Verification Service Class. Upon receipt from an SCU of a verification of communication request, the Imager will issue confirmation.

3.2.3.2 Delivery- Basic Print Management SCP

The Imager provides standard SCP conformance to the DICOM Basic Print Management SOP Class. Association attempts will be rejected if more than the maximum number of simultaneous delivery SCP associations is attempted.

3.3 Basic Print Management Meta SOP Class

The Meta SOP Class is defined by the following set of supported SOP Classes:

SOP Class	UID Value
Basic Film Session SOP Class	1.2.840.10008.5.1.1.1
Basic Film Box SOP Class	1.2.840.10008.5.1.1.2
Basic Grayscale Image Box SOP Class	1.2.840.10008.5.1.1.4
Basic Color Image Box SOP Class	1.2.840.10008.5.1.1.4.1
Printer SOP Class	1.2.840.10008.5.1.1.16

SCUs should not mix color and grayscale images at the film session level. They should create separate color and grayscale film sessions in succession.

Note: In the sections below,, the customer user interface provides the ability to configure the attributes and default values.

3.3.1 Basic Film Session SOP Class

3.3.1.1 DIMSE Service N-CREATE

Attribute	SCP Usage	Tag	Possible Values	Default Values
Number of Copies	U	(2000,0010)	1-99	1
Print Priority	U	(2000,0020)	HIGH, MED, LOW	MED, Configurable from the Imager's User Interface.
Medium Type	U	(2000,0030)	BLUE FILM, PAPER	BLUE FILM Configurable from the Imager's User Interface.
Film Destination	U	(2000,0040)	All values are converted to BIN_1.	BIN_1
Film Session Label	U	(2000,0050)	Up to 64 characters may be provided	Null String
Memory Allocation	U	(2000,0060)	Not used	Not used

3.3.1.2 Film Session N-Create Status Code

Code	Status	Action/Meaning
0000H	Success	Film session created. Some attributes may have different values than those that were requested. The changed attributes will be returned with the values that were used. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0210H	Failure	A film session already exists. Another is not created. (DICOM PS 3.7 Annex C.5.9)

3.3.1.3 DIMSE Service N-Action

The Imager uses N-ACTION to accept print commands from the SCU. Once a print command is received, the Imager prints all films in the session. The Imager conforms to the N-ACTION specification in PS 3.4 Annex H.4.1.2.4 of the DICOM standard, and the Imager collates all film boxes when printed.

3.3.1.4 Film Session N-Action Status Code

Code	Status	Action/Meaning
0000H	Success	All images in the session are printed as specified. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0112H	Failure	The SOP Instance UID of the requested film session is returned. (DICOM PS 3.7 Annex C.5.19)

B602H	Warning	Nothing is printed. (DICOM PS 3.4 Annex H.4.1.2.4.2)
C600H	Failure	Nothing is printed. (DICOM PS 3.4 Annex H.4.1.2.4.2)

3.3.1.5 DIMSE Service N-SET

The Imager uses N-SET to update the Film Session values as supplied by the SCU.

3.3.1.6 Film Session N-Set Status Code

Code	Status	Action/Meaning
0000H	Success	Film session data is set. Some attributes may have different values than those that were requested. The changed attributes will be returned with the values that were used. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0112H	Failure	The SOP Instance UID of the requested film session is returned. (DICOM PS 3.7 Annex C.5.19)

3.3.1.7 Film Session N-Delete Status Code

Code	Status	Action/Meaning
0000H	Success	The Film session is deleted. (DICOM PS 3.7 Annex C.1.1)
0112H	Failure	The SOP Instance UID of the specified film session was not found. (DICOM PS 3.7 Annex C.5.19)

3.3.2 Basic Film Box SOP Class

3.3.2.1 DIMSE Service N-CREATE

Attribute	SCP Usage	Tag	Possible Values	Default Values
Image Display Format	M	(2010,0010)	<p>STANDARD\C,R For both PORTRAIT and LANDSCAPE Film Orientation, where C=1-9 and R=1-9 (i.e. support 1-up, 2-up, 4-up, 6-up, 8-up, 9-up, 12-up, 15-up, 16-up, 20-up, 24-up, 30-up, 35-up, 42-up standard formats)</p> <p>ROWr1,r2,r3...where r1, r2, r3...is the number of images in each row.</p> <p>The rows are limited to 10 and the number of images in each row is limited to 10.</p> <p>CUSTOMI I = 101, 102 Only valid for PORTRAIT Film Orientation (2010,0040).</p> <p>Notes: See Annex B for description</p>	None. SCU must provide.
Referenced Film Session Sequence	M	(2010,0500)		None. SCU must provide.
>Referenced SOP Class UID	M	(0008,1150)		None. SCU must provide.
>Referenced SOP Instance UID	M	(0008,1155)		None. SCU must provide.
Referenced Basic Image Box Sequence	M	(2010,0510)		None. SCU must provide.
Referenced Basic Annotation Box Sequence	MC	(2010,0520)		None.
Film Orientation	U	(2010,0040)	PORTRAIT, LANDSCAPE	PORTRAIT

Attribute	SCP Usage	Tag	Possible Values	Default Values
Film Size ID	U	(2010,0050)	8INX10IN 10INX12IN 11INX14IN 14INX17IN 8_5INX11IN A3 A4 See Annex C for detailed explanation	14"X17" or The largest size supported by the printer. Configurable from the Imager's User Interface. First four sizes listed are Blue Film only. Last three sizes listed are Paper only.
Magnification Type	U	(2010,0060)	REPLICATE, BILINEAR, CUBIC, NONE	CUBIC Configurable from the Imager's User Interface.
Image Max Density	U	(2010,0130)	0-399** **Note: This range is supported for compatibility. The actual achievable density range will be determined by the system capability.	280 Configurable from the Imager's User Interface.
Configuration Information	U	(2010,0150)	Curve Shape (CS): 000 to 999 Contrast Values (CN): -1 to -5 Lower contrast 0 Normal +1 to +5 Higher contrast Pivot Density (PD): 0 to 2.4 in increments of 0.2 Perception LUT Selection (LUT): LUT=m, n (m=string, n = 1 to 15) For TFT n = 1 to 15 For ULUT n = 1 to 12 Text Macros (TM): %PRNTDAT%, %TIM%, %FOF%, %\$TIME\$, %SES% Perception LUT cannot be used with Curve Shape, Contrast or Pivot Density. See Annex A for description	Configurable from the Imager's User Interface. Configurable from the Imager's User Interface. 1.2 LUT=Ver693c0.w87,6 None

Attribute	SCP Usage	Tag	Possible Values	Default Values
Referenced Presentation LUT Sequence	MC	(2050,0500)	If the PLUT is received, the tonescaling data from the above Configuration Information will be ignored.	None
>SOP Class UID	MC	(0008,1150)		None
>SOP Instance UID	MC	(0008,1155)		None
Annotation Display Format ID	U	(2010,0030)	<p>0 - No annotation</p> <p>1 - Text centered at bottom of film</p> <p>6 – Six annotation positions on two lines, centered at bottom of film.</p> <p>NONE – No annotation</p> <p>LABEL – Annotation at bottom of film.</p> <p>BOTTOM – Text at bottom of images.</p> <p>COMBINED – 1 line at the bottom of the page and 1 line under each image.</p> <p>See Basic Annotation Box SOP (section 2.5) for valid values for Annotation Position (2030,0010) for each of these formats.</p>	0 (No annotation)
Smoothing Type	U	(2010,0080)	<p>NORMAL (minimum cubic convolution error)</p> <p>ENHANCED, ENHANCED1 (Valid only for Magnification Type CUBIC.)</p> <p>0-15 (Valid only for Magnification Type CUBIC.)</p> <p>If Smoothing Type is unrecognized, NORMAL will be used.</p>	<p>5</p> <p>Configurable from the Imager's User Interface.</p>
Border Density	U	(2010,0100)	<p>BLACK, WHITE, i</p> <p>where i may = 0 -399**</p> <p>**Note: This range is supported for compatibility. The actual achievable density range will be determined by the system capability.</p>	<p>Image Max Density</p> <p>Configurable from the Imager's User Interface.</p>

Attribute	SCP Usage	Tag	Possible Values	Default Values
Min Density	U	(2010,0120)	0-399** (Value must be less than Max Density [2010,0130]) This value is used only when PLUT or Curve Shape is applied to the images on the page. When PLUT or Curve Shape is applied, actual minimum density will be the greater of the user requested value and the Dmin of the film. **Note: This range is supported for compatibility. The actual achievable density range will be determined by the system capability.	Dmin of the Film
Illumination	MC	(2010,015E)	Positive integer in units of cd/m2	2000 Configurable from the Imager's User Interface.
Reflective Ambient Light	MC	(2010,0160)	Positive integer in units of cd/m2	10 Configurable from the Imager's User Interface.
Trim	U	(2010,0140)	YES, NO	NO Configurable from the Imager's User Interface.

3.3.2.2 Film Box N-Create Status Code

Code	Status	Action/Meaning
0000H	Success	Film box created. Some attributes may have different values than those that were requested. The changed attributes will be returned with the values that were used. (DICOM PS 3.7 Annex C.1.1)
0106H	Failure	The invalid attribute value is returned in the response data set. (DICOM PS 3.7 Annex C.5.11)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0112H	Failure	The film session requested to contain this film box does not exist. (DICOM PS 3.7 Annex C.5.19)
0213H	Failure	Page limit is exceeded. (DICOM PS 3.7 Annex C.5.22)
0120H	Failure	The attribute tag of the missing required attribute is returned. (DICOM PS 3.7 Annex C.5.13)
0121H	Failure	The film box is not created. The required attribute was present, but contained no value. (DICOM PS 3.7 Annex C.5.13)

3.3.2.3 DIMSE Service N-ACTION

The Imager uses the N-ACTION to accept print instruction from the SCU. When such an instruction is received, the Imager prints the current film boxes in the session.

3.3.2.4 FilmBox N-Action Status Code

Code	Status	Action/Meaning
0000H	Success	All images in the film box are printed as specified. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0112H	Failure	The film box does not exist. (DICOM PS 3.7 Annex C.5.19)
B603H	Warning	Nothing is printed (DICOM PS 3.4 Annex H.4.1.2.42)
C600H	Failure	Nothing is printed. (DICOM PS 3.4 Annex H.4.1.2.4.2)

3.3.2.5 DIMSE Service N-SET

The Imager uses N-SET to update the Basic Film Box values as supplied by the SCU. The following attributes may be updated:

Attribute	SCP Usage	Tag
Magnification Type	U	(2010,0060)
Max Density	U	(2010,0130)
Configuration Information	U	(2010,0150)
Smoothing Type	U	(2010,0080)
Border Density	U	(2010,0100)
Min Density	U	(2010,0120)
Illumination	MC	(2010,015E)
Reflective Ambient Light	MC	(2010,0160)
Trim	U	(2010,0140)

3.3.2.6 FilmBox N-Set Status Code

Code	Status	Action/Meaning
0000H	Success	Film box data is set. Some attributes may have different values than those that were requested. The changed attributes will be returned with the values that were used. (DICOM PS 3.7 Annex C.1.1)
0106H	Failure	The invalid attribute value is returned in the response data set. (DICOM PS 3.7 Annex C.5.11)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0112H	Failure	The specified film box does not exist. (DICOM PS 3.7 Annex C.5.19)
0120H	Failure	The attribute tag of the missing required attribute is returned. (DICOM PS 3.7 Annex C.5.13)
0121H	Failure	The required attribute was present, but contained no value. (DICOM PS 3.7 Annex C.5.13)
0213H	Failure	Page limit is exceeded. (DICOM PS 3.7 Annex C.5.22)

3.3.2.7 DIMSE Service N-DELETE

Upon receipt of an N-DELETE from the SCU, the Imager removes the individual image boxes from the session.

3.3.2.8 FilmBox N-Delete Status Code

Code	Status	Action/Meaning
0000H	Success	The film box is deleted. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0112H	Failure	The SOP Instance UID of the specified film session is returned. (DICOM PS 3.7 Annex C.5.19)

3.3.3 Basic Grayscale Image Box SOP Class

3.3.3.1 DIMSE Service N-SET

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
Image Position	M	(2020,0010)	All values within the range of Image Display Format	None. SCU must provide.
Preformatted Grayscale Image Sequence	M	(2020,0110)		None. SCU must provide.
>Samples Per Pixel	U	(0028,0002)	1	1
>Photometric Interpretation	U	(0028,0004)	MONOCHROME1, MONOCHROME2	MONOCHROME2
>Rows	M	(0028,0010)	Maximum Values: Depends on film size and printer model. The aspect ratio is used with the printer's page extents, display format, etc. to calculate this value. See Annex D.	None. SCU must provide.
>Columns	M	(0028,0011)	Maximum Values: Depends on film size and printer model. The aspect ratio is used with the printer's page extents, display format, etc. to calculate this value. See Annex D.	None. SCU must provide.
>Pixel Aspect Ratio	MC	(0028,0034)	R\C R, C = 1 to 9999 (Integer)	1\1
>Bits Allocated	M	(0028,0100)	8, 16	None. SCU must provide.

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
>Bits Stored	M	(0028,0101)	8, 10, 12, 14	None. SCU must provide.
>High Bit	M	(0028,0102)	Bits Stored -1	None. SCU must provide.
>Pixel Representation	M	(0028,0103)	0000H (unsigned integer)	0000H
>Pixel Data	M	(7FE0,0010)	All values consistent with Bits Stored	None. SCU must provide.
Polarity	U	(2020,0020)	NORMAL, REVERSE	NORMAL
Magnification Type	U	(2010,0060)	REPLICATE, BILINEAR, CUBIC, NONE	CUBIC Configurable from the Imager's User Interface.
Smoothing Type	U	(2010,0080)	NORMAL (minimum cubic convolution error) ENHANCED, ENHANCED1 (Valid only for Magnification Type CUBIC.) 0-15 (Valid only for Magnification Type CUBIC.)	5
Configuration Information	U	(2010,0150)	Setting these values will override film box settings for this image position. Curve Shape (CS): 000 to 999 Perception LUT Selection (LUT): LUT=m, n (m=string, n = 0 to 15) Curve Shape and Perception LUT are mutually exclusive. Note that m can be 0. See Annex A for description	Configurable from the Imager's User Interface. Configurable from the Imager's User Interface. <i>LUT=Ver693c0.w87,6</i>
Requested Image Size	U	(2020,0030)	Width of Image Box in millimeters (fractional millimeters supported) 0.00 indicates, "Maximize film utilization while maintaining Image aspect ratio". If this value exceeds the available dimensions of the Image Box, it will be accepted only if the Requested Decimate/Crop Behavior value is NOT set to FAIL.	0.00
Requested Decimate/Crop Behavior	U	(2020,0040)	DECIMATE/CROP/FAIL DECIMATE: If the Image Size exceeds the printable area, the Image Size will be	DECIMATE Configurable from the Imager's User Interface.

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
			<p>reduced while preserving the full view of the Image.</p> <p>CROP: If the Image Size exceeds the printable area, the Image will be center cropped by removing pixels that fall outside the printable area.</p> <p>FAIL: If the Image Size exceeds the printable area, the Image will be rejected.</p>	
Referenced Presentation LUT Sequence	U	(2050,0500)	If the PLUT is received, the tonescaling data from the above Configuration Information will be ignored	

3.3.3.2 ImageBox N-Set Status Code

Code	Status	Action/Meaning
0000H	Success	Image box data is set. Some attributes may have different values than those that were requested. The changed attributes will be returned with the values that were used. (DICOM PS 3.7 Annex C.1.1)
0106H	Failure	The invalid attribute value is returned in the response data set. (DICOM PS 3.7 Annex C.5.11)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)
0112H	Failure	The specified film box does not exist. (DICOM PS 3.7 Annex C.5.19)
0120H	Failure	The attribute tag of the missing required attribute is returned. (DICOM PS 3.7 Annex C.5.13)
0121H	Failure	The required attribute was present, but contained no value. (DICOM PS 3.7 Annex C.5.13)
0213H	Failure	Page limit is exceeded. (DICOM PS 3.7 Annex C.5.22)
C603H	Failure	Image Size cannot be achieved with the requested decimate crop behavior.

3.3.4 Basic Color Image Box SOP Class

3.3.4.1 DIMSE Service N-SET

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
Image Position	M	(2020,0010)	All values within the range of Image Display Format	None. SCU must provide.
Basic Color Image Sequence	M	(2020,0111)		

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
>Samples Per Pixel	U	(0028,0002)	3	3
>Photometric Interpretation	U	(0028,0004)	RGB	RGB
>Planar Configuration	M	(0028,0006)	000H or 001H 000H – pixels arrive in R1G1B1R2G2B2R3G3B3... order 001H – pixels arrive in R1R2R3..., G1G2G3... B1B2B3... order	None. SCU must provide.
>Rows	M	(0028,0010)	See Annex D for behavior depending on the printed image size.	None. SCU must provide.
>Columns	M	(0028,0011)	See Annex D for behavior depending on the printed image size.	None. SCU must provide.
>Pixel Aspect Ratio	MC	(0028,0034)	R\C R, C = 1 to 9999 (Integer)	1\1
>Bits Allocated	U	(0028,0100)	8	8
>Bits Stored	U	(0028,0101)	8	8
>High Bit	U	(0028,0102)	7	7
>Pixel Representation	M	(0028,0103)	0000H (unsigned integer)	0000H
>Pixel Data	M	(7FE0,0010)	All values consistent with Bits Stored	None. SCU must provide.
Polarity	U	(2020,0020)	NORMAL, REVERSE	NORMAL
Magnification Type	U	(2010,0060)	REPLICATE, BILINEAR, CUBIC, NONE All values are used only when the printer cannot print color. When the printer prints color, REPLICATE is always used.	CUBIC Configurable from the Imager's User Interface.
Smoothing Type	U	(2010,0080)	NORMAL (minimum cubic convolution error) ENHANCED, ENHANCED1 (Valid only for Magnification Type CUBIC.) 0-15 (Valid only for Magnification Type CUBIC.) Smoothing Type is used only when the printer cannot print color.	5

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
Configuration Information	U	(2010,0150)	<p>Setting these values will override film box settings for this image position.</p> <p>Curve Shape (CS): 000 to 999</p> <p>Perception LUT Selection (LUT): LUT = m, n (m=string, n = 0 to 15)</p> <p>Curve Shape and Perception LUT are mutually exclusive.</p> <p>Curve Shape or Perception LUT is used only when the printer cannot print color.</p> <p>Note that m can be 0. See Annex A for description</p>	<p>Configurable from the Imager's User Interface.</p> <p>Configurable from the Imager's User Interface.</p> <p>LUT=Ver693c0.w87,6</p>
Requested Image Size	U	(2020,0030)	<p>Width of Image Box in millimeters (fractional millimeters supported)</p> <p>0.00 indicates, "Maximize film utilization while maintaining Image aspect ratio".</p> <p>If this value exceeds the available dimensions of the Image Box, it will be accepted only if the Requested Decimate/Crop Behavior value is NOT set to FAIL.</p>	0.00
Requested Decimate/Crop Behavior	U	(2020,0040)	<p>DECIMATE/CROP/FAIL</p> <p>DECIMATE: If the Image Size exceeds the printable area, the Image Size will be reduced while preserving the full view of the Image.</p> <p>CROP: If the Image Size exceeds the printable area, the Image will be center cropped by removing pixels that fall outside the printable area.</p> <p>FAIL: If the Image Size exceeds the printable area, the Image will be rejected.</p>	<p>DECIMATE</p> <p>Configurable from the Imager's User Interface.</p>

3.3.4.2 Basic Color Image Box N-Set Status Code

Code	Status	Action/Meaning
0000H	Success	Image box data is set. Some attributes may have different values than those that were requested. The changed attributes will be returned with the values that were used. (DICOM PS 3.7 Annex C.1.1)
0106H	Failure	The invalid attribute value is returned in the response data set. (DICOM PS 3.7 Annex C.5.11)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex

		C.5.21)
0112H	Failure	The specified film box does not exist. (DICOM PS 3.7 Annex C.5.19)
0120H	Failure	The attribute tag of the missing required attribute is returned. (DICOM PS 3.7 Annex C.5.13)
0121H	Failure	The required attribute was present, but contained no value. (DICOM PS 3.7 Annex C.5.13)
0213H	Failure	Page limit is exceeded. (DICOM PS 3.7 Annex C.5.22)

3.3.5 Printer SOP Class

3.3.5.1 DIMSE Service N-GET

Changes in printer status will be sent when they occur using N-EVENT-REPORT only if the SCU established the association using a called AE title beginning with “NER_”. Otherwise, the SCU can use the N-GET to retrieve an instance of the Printer SOP class.

Printer Status (2110,0010) and Printer Status Info (2110,0020) will be returned with all N-GET requests of the Printer SOP class.

Attribute	SCP Usage	Tag	Supported Values
Printer Status	M	(2110,0010)	NORMAL,WARNING, FAILURE
Printer Status Info	M	(2110,0020)	for Printer Status of NORMAL: NORMAL
			for Printer Status of WARNING: CHECK PRINTER EMPTY 8X10 BLUE EMPTY 10X12 BLUE EMPTY 11X14 BLUE EMPTY 14X17 BLUE EMPTY 8.5X11 PAPER EMPTY A3 PAPER EMPTY A4 PAPER COVER OPEN FILM JAM BAD SUPPLY MAGAZINE PRINTER OFFLINE
			for Printer Status of FAILURE: PRINTER DOWN
Printer Name	U	(2110,0030)	Any value up to 16 characters in length. Chosen by user at time of installation

Attribute	SCP Usage	Tag	Supported Values
Printer Manufacturer	U	(0008,0070)	Carestream Health
Printer Manufacturer Model Name	U	(0008,1090)	DVChroma
Printer Device Serial Number	U	(0018,1000)	AAAAAAAA (number up to 8 ASCII characters)
Software Version	U	(0018,1020)	The actual Imager software version (ID up to 6 ASCII characters) as X.y where X is the Imager Release number and y the version number.
Date of Last Calibration	U	(0018,1200)	Not supported
Time of Last Calibration	U	(0018,1201)	Not supported

3.3.5.2 Printer SOP N-Get Status Code

Code	Status	Action/Meaning
0000H	Success	Printer Status and Printer Status Info are always returned along with the requested attribute values. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)

3.4 Basic Annotation Box SOP Class

3.4.1 DIMSE Service N-SET

The Basic Annotation Box SOP Instance is created by the SCP at the time of the Basic Film Box SOP Instance is created, based on the value of the Annotation Display Format ID attribute (2010,0030) of the Basic Film Box.

A maximum of 6 annotation boxes may be received with a film box.

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
Annotation Position	M	(2030,0010)	<p>Annotations are placed in order from upper-left to lower right.</p> <p>If the Annotation Display Format ID is 1, then value must be 1. The text will be printed on one line at the bottom of the film.</p> <p>If the Annotation Display Format ID is 6, then valid range is 1-6. The text will be printed within 2 lines at the bottom of the film, within 6 different positions.</p> <p>If the Annotation Display Format ID is LABEL, the valid range is 0-1. The text will be printed at the bottom of the film on two lines.</p> <p>If the Annotation Display Format ID is BOTTOM, then the valid range is 1 to the number of images in the Film Box. The text will be placed below the images.</p> <p>If the Annotation Display Format ID is COMBINED, then the valid range is 0 to the number of images in the Film Box. Position 0 will be printed at the bottom of the film. The other annotations will be printed below the images.</p> <p>Any annotation box with a position outside the valid range will be ignored.</p>	None.
Text String	M	(2030,0020)	Up to 64 characters (see Note)	None.

Note: The number of characters displayed may be less than 64 characters depending on the size of the film, the page format, the annotation format and the characters used. This exception does not apply to Annotation Format IDs of LABEL or 1.

3.4.2 Annotation N-Set Status Code

Code	Status	Action/Meaning
0000H	Success	The annotation data is set.
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.1.1)
0112H	Failure	The annotation box does not exist. (DICOM PS 3.7 Annex C.5.21)
0116H	Warning	Invalid Position was specified. (DICOM PS 3.7 Annex C.5.19)
0213H	Failure	Page limit is exceeded. (DICOM PS 3.7 Annex C.5.22)

3.5 Presentation LUT SOP Class

3.5.1 DIMSE Service N-CREATE

The Presentation LUT SOP Instance is created by the SCP upon receipt of the N-CREATE action. The Print SCU may create Presentation LUT instance prior to being referenced by the Basic Film Box. Multiple Presentation LUT instances are supported in an association, but only one instance will be supported for each image.

The SCU shall send either Presentation LUT Sequence or the Presentation LUT Shape. These values are mutually exclusive and the action will result in an error if neither or both are present. The presence of the Presentation LUT instance overrides any data set in the Configuration Information attribute (2010,0150) of the Film Box or Image Box.

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
Presentation LUT Sequence	M	(2050,0010)		None.
>LUT Descriptor	M	(0028,3002)	The first value is the number of entries in the lookup table. The number of entries shall be equal to the number of possible values in the input. (For 8 bit input will be 256 entries, for 12 bit input it will be 4096 entries) The second value is the first input value mapped, and shall always be 0. The third value specifies the number of bits for each entry in the LUT Data. It shall be between 10 and 16 inclusive.	None.
>LUT Explanation	U	(0028,3003)	Free form text explanation of the meaning of the LUT.	None.

Attribute & Usage	SCP Usage	Tag	Supported Values	Default Values
>LUT Data	M	(0028,3006)	The LUT Data shall be stored in a format equivalent to 16 bits allocated where the high bit is equal to bits stored - 1, where bits stored is the third value of the LUT Descriptor.	None.
Presentation LUT Shape	M	(2050,0020)	Enumerated values IDENTITY and LIN OD.	None.

3.5.2 Presentation LUT N-Create Status Code.

Code	Status	Action/Meaning
0000H	Success	The Presentation LUT is created. Some attributes may have different values than those that were requested. The changed attributes will be returned with the values that were used. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)

3.5.3 DIMSE Service N-DELETE

Upon receipt of an N-DELETE from the SCU, the Imager removes the Presentation LUT instance.

3.5.4 Presentation LUT N-Delete Status Code

Code	Status	Action/Meaning
0000H	Success	The Presentation LUT is deleted. (DICOM PS 3.7 Annex C.1.1)
0110H	Failure	The Error Comment field will contain the detailed description of the error. (DICOM PS 3.7 Annex C.5.21)

4 Communication Profiles

4.1 Supported Communications Stacks

The Imager provides TCP/IP Network Communication Support as defined in Part 8 of the DICOM standard.

4.2 Physical Media

The Imager supports *Ethernet* with the following physical connections:

Standard Twisted pair (10BaseT, 100BaseT, and 1000BaseT)

5 Extensions/Specializations/Privatizations

There are not any extensions/specializations/privatizations.

6 Configuration

The following attributes are configurable by the user:

- IP address
- DICOM Port number
- Subnet Mask
- Local Network Host Name (Imager's AE Title)
- Network Gateway
- DNS Server IP
- Number of maximum associations
- DICOM Service(s) available
- Basic Color Print Management Association
- N-Event-Report support (Default is Off)
- Other configuration indicated in the SOP Class tables.
- Other properties as indicated in User Guide for this product.

7 Support of Extended Character Sets

The Imager supports the ISO-IR 100 Latin 1 character set as well as the ISO-IR 6 default character set.

The Imager also supports the ISO-IR 87 character set. This is part of the JIS X 0208 code table for 2-byte Japanese character sets that supports Kanji (ideograph), Hiragana (phonetic), and Katakana (phonetic).

The Imager also supports the ISO-IR 13 character set. This is part of the JIS X 0201 code table for single-byte Japanese Katakana (phonetic) characters.

The value set in the tag Specific Character Set (0008,0005) must be either:

1. “ISO_IR 6” (default repertoire)
2. “ISO_IR 100” (Latin 1)
3. “ISO_IR 13” (Katakana)
4. “ISO 2022 IR 13\ISO 2022 IR 87” (Katakana, Hiragana, Kanji)
5. “ISO 2022 IR 159” (supplementary Kanji set)
6. “ISO 2022 IR 149” (Korean Hangu set)
7. “GB18030” (Chinese set)
8. or blank (ISO-IR 6 is the default character set)

8 Error Handling

Warnings indicate that the operation/notification has been completed, but an error was detected. Under warning conditions, operations continue forward without additional actions.

Failures convey that the operation/notification failed and was not performed. Refer to the DICOM Specification PS 3.7, Annex C for the Status Types supported by the DIMSE services.

Annex A - Configuration Information

The Configuration Information attribute contains the list of specific values. These attributes are not DICOM standard attributes.

The Configuration Information value is an ordered list. The attribute is specified using the ASCII two-character key prefix in the following sequence:

1. Curve Shape, Contrast, Pivot Density
or Perception LUT
2. Text Macros.

The Film Box Curve Shape value applies to all images in the Film Box except when Curve Shape or Perception LUT is specified for the image in the Image Box.

Attribute	Usage	Description	Default
Curve Shape designated by the ASCII two-character prefix: CS	U/M	000 to 999 Note: 000 = linear, 999 = highest curvature Curve Shape is a tone scale adjustment used to optimize the image on film compared to the image on the operator console monitor. Curve shape is not valid when a Perception LUT is specified.	Film Box: Value set in the Printer by the user Image Box: Basic Film Box Curve Shape
Contrast designated by the ASCII two-character prefix: CN	U/M	-5 to 5 Note: Integer values only. Negative Contrast settings are lower contrast where the amount of data that is represented by medium film densities is increased. Positive settings are higher contrast where the amount of data that is represented by high and low densities is increased.	Value set in the Printer by the user
Pivot Density designated by the ASCII two-character prefix: PD	U/M	0.0 to 2.4 Note: Value must be specified in increments of 0.2. Densities above and below the pivot density will be adjusted up and down by an amount that is a function of the difference between the code value and the pivot density code value.	Value set in the Printer by the user
Perception LUT Selection designated by the ASCII three-character prefix: LUT	U/M	LUT = m, n Allows selection of the LUT that best suits the user's images. M is the name of the TFT set and N specifies a contrast setting within the group. Curve shape will be ignored if the LUT parameter is used. m=string (0=default group) n=0 to 15 (0=use default value set at User Interface)	m=0, n=0

Attribute	Usage	Description	Default
Text Macros designated by the ASCII two-character key prefix: TM	M/M	%PRNTDAT% Capture Date* DD-MMM-YY %TIM% Capture Time* HH:MM (HH=0-23) %FOF% Film of Film Count NN/MM %\$TIME\$% Time of Printing HH:MM:SS %SES% Film Session Label AAAAAA (1-64 chars from the Film Session SOP Class) Note: The text macros will be printed on the bottom of the film and will be truncated if necessary	None

*** This is the receiving Date/Time of the first valid image of the first Film Box.**

Examples:

"CS333"

The curve shape is set to 1/3 of the printer's tone scale range and defaults are applied to contrast and pivot density.

"CS500\CN3\PD2.2"

The curve shape is set to 1/2 the printer's tone scale range, Contrast is set to 3, and pivot density is set to 2.2.

"PD2.0"

The pivot density is set to 2.0, and defaults are applied to curve shape and contrast.

"CS333\CN3\PD2.2\TM%PRNTDAT%%TIM%%FOF%"

The curve shape is set to 1/3 of the range, Contrast is set to 3, and pivot density is set to 2.2.

The following text macros will be printed on the bottom of the page:

Date of Printing, Time of Printing, and Film of Film count.

"LUT=Ver693c0.w87,3"

The Perception LUT TFT set is "Ver693c0.w87" and the Contrast Setting is 3.

"LUT=0,3\ TM%PRNTDAT%%TIM%%FOF%"

The Perception LUT TFT set is 0 (default) and the Contrast Setting is 3.

The following text macros will be printed on the bottom of the page:

Date of Printing, Time of Printing, and Film of Film count.

"TM%PRNTDAT%%TIM%%FOF%"

The following text macros will be printed at the bottom of the page:

Date of Printing, Time of Printing, and Film of Film count.

"PD2.0\CN4\CS333"

This is **invalid** because the attributes are out of order, curve shape must precede pivot density and contrast, and contrast must precede pivot density. It should be "CS333\CN4\PD2.0".

"CS333\PD1.2\LUT=0,3"

This is **invalid** because Curve Shape and Pivot Density cannot be mixed with Perception LUT. In this case, the Perception LUT setting will be used.

Annex B - Custom Formats


The following formats are expected to be supported by imagers manufactured by Carestream Health and will only be specific to these printers.

Format ID 101

This format consists of 7 image positions, 1 large image in the upper section of the page and 6 smaller images in the lower section of the page. The size and positioning of the images are defined in terms of the standard formats 2 and 12.

Upper Section: 1 frame of a 2-up format.

Lower Section: 6 frames of a 12-up format.




Format ID 102

This format consists of 11 image positions, 2 large images in the upper section of the page and 9 smaller images in the lower section of the page. The size and positioning of the images are defined in terms of the standard formats 6 and 15.

Upper Section: 2 frames occupying top 1/3 of media.

Lower Section: 9 frames occupying bottom 2/3 of media.



Annex C - Unsupported Film Types

The following table illustrates what happens when the Imager receives an unsupported Film Size ID [DICOM element (2010,0050)] or Medium Type ID (2000,0030) from an SCU. This section is added to clarify the Imager's behavior when these conditions occur.

The Imager initializes a list with the Media Type enumerations for known reflective medias and keeps a collection of media Types that do not allow substitution. For the Type that allows substitution, this priority is followed:

1. The highest priority is placed on matching the film base's reflective property.
2. The next priority is granted to the Film Size ID (2010,0050). The Imager will attempt to match the film size to that requested in the N-CREATE Film Box.

Note: For best results, Print SCUs should either specify films that are installed in the imager or not specify these attributes.

The following tables list the details unique to specific printer model that recognizes the Medium Type currently installed.

Medium Type Supported	Medium Type Currently Installed	Size Supported	Size Currently Installed	Result
Yes	Yes	Yes	Yes	The imager prints the page on the specified film type.
Yes	Yes	Yes	No	The imager stores the page in a queue. User must change film magazines to appropriate size and base to get printout.
Yes	Yes	No	N/A	The Imager replaces the SCU's Film Size of the same base selection with the next larger supported size for the imager. This new Film Size ID is what is passed to the printer. If no larger film size is available, then the next smaller supported film size is selected. In the Imager's response to the SCU, the Film Size ID (2010,0050) is modified so that it contains the new Film Size ID. The page is either queued in the imager or printed.
Yes	No	Yes	Yes	The Imager replaces the SCU's Medium Type selection with the supported medium type for the selected film size for the imager. The imager prints the page on the specified film type.
Yes	No	No	N/A	The Imager replaces the SCU's Film Size selection with the next larger supported size of the same base for the imager. If no larger film size is available, then the next smaller supported

Medium Type Supported	Medium Type Currently Installed	Size Supported	Size Currently Installed	Result
				film size of the same base is selected. This new Film Size ID is what is passed to the imager. In the Imager's response to the SCU, the Film Size ID (2010,0050) is modified so that it contains the new Film Size ID. The page is queued in the imager, and printed when the film size and base is available.
No	N/A	Yes	No	The Imager replaces the SCU's Medium Type selection with the supported medium type for the selected film size for the imager. This new Medium Type ID is what is passed to the imager. The page is queued in the imager.
No	N/A	No	N/A	The Imager replaces the SCU's Film Size selection with next larger supported size for the destination. This new Film Size ID is what is passed to the imager. The medium type of this new film size will also be selected. In the Imager's response to the SCU, the Film Size ID (2010,0050) is modified so that it contains the new Film Size ID. The page is queued in the imager.

Annex D – Printer Specifications

The following tables list the details unique to specific the Imager.

D.1 Medium Size

This table shows the accepted film and paper sizes. The maximum image size represents the largest image accepted with and without annotation printed on the bottom of the film.

If annotations are printed, some pixel space will be used for the text and the maximum image size will be reduced.

If the requested image size is smaller than the image size, then minification option is needed to get the image printed.

Portrait and Landscape Size Info:

Available Sizes	Max Width (Portrait) (Pixel pitch =14.17)		Max Height w/o Annotation (Portrait)	Max Height w/ Annotation (Portrait)
	Pixels	Mm		
8INX10IN	2760	194.78	3300	3052
10INX12IN	3484	245.87	4016	3768
11INX14IN	3862	272.55	4732	4484
14INX17IN	4916	346.93	5810	5562
8.5INX11IN	2974	209.88	3854	3606
A3	4122	290.90	5852	5604
A4	2890	203.95	4108	3860
Available Sizes	Max Width (Landscape) (Pixel pitch = 14.17)		Max Height w/o Annotation (Landscape)	Max Height w/ Annotation (Landscape)
	Pixels	Mm		
8INX10IN	3300	232.89	2760	2512
10INX12IN	4016	283.42	3484	3236
11INX14IN	4732	333.94	3862	3614
14INX17IN	5810	410.02	4916	4668
8.5INX11IN	3854	271.98	2974	2726
A3	5852	412.99	4122	3874
A4	4108	289.91	2890	2642

NOTE:

Film is only available in the following sizes: 8x10, 10x12, 11x14, and 14x17.

Paper is only available in the following sizes: 8.5x11, A3 and A4.

The above tables implicitly lists the maximum image values of a 1Up given the DICOM Requested Decimate/Crop Behavior attribute set to FAIL. For all other formats the maximum image values follows these general rules:

- In any Row, the sum of all Image Widths plus all Horizontal Separations cannot exceed the above Max Width values.
- In any Column, the sum of all Image Lengths plus all the Vertical Separations cannot exceed the above Max Height values.

D.2 Printer Capabilities

The following table shows the Printable Max Area for a sample of Multiple Page Formats when horizontal and vertical separations are set to zero.

Format	8X10		10X12		11X14		14X17		8.5X11		A3		A4	
	Width	Height	Width	Height	Width	Height	Width	Height	Width	Height	Width	Height	Width	Height
1	2760	3300	3484	4016	3862	4732	4916	5810	2974	3854	4122	5852	2890	4108
2	2760	1650	3484	2008	3862	2366	4916	2905	2974	1927	4122	2926	2890	2054
4	1380	1650	1742	2008	1931	2366	2458	2905	1487	1927	2061	2926	1445	2054
6	1380	1100	1742	1338	1931	1577	2458	1936	1487	1284	2061	1950	1445	1369
8	1380	825	1742	1004	1931	1183	2458	1452	1487	963	2061	1463	1445	1027
9	920	1100	1161	1338	1287	1577	1638	1936	991	1284	1374	1950	963	1369
12	920	825	1161	1004	1287	1183	1638	1452	991	963	1374	1463	963	1027
15	920	660	1161	803	1287	946	1638	1162	991	770	1374	1170	963	821
16	690	825	871	1004	965	1183	1229	1452	743	963	1030	1463	722	1027
20	690	660	871	803	965	946	1229	1162	743	770	1030	1170	722	821
24	690	550	871	669	965	788	1229	968	743	642	1030	975	722	684
30	552	550	696	669	772	788	983	968	594	642	824	975	578	684
35	552	471	696	573	772	676	983	830	594	550	824	836	578	586
42	460	471	580	573	643	676	819	830	495	550	687	836	481	586

Note: This assumes no page annotation and in portrait mode. For landscape, the width and height would be reversed.